

6 out of 6 2dvds are calibrated and operating

Submitted by Patrick.Gatlin on Mon, 2013-04-29 22:03

Date:

Monday, April 29, 2013 - 22:00

Instrument List:

- [Green](#)

Discussion:

All 2dvds have been calibrated as of Monday (4/29). Remote connection to sn25, 35 and 70 have been reliable. However the remote connections to sn36, 37 and 38 have not been very stable. We are unsure if this is a result of poor signal quality on the Verizon cellular network or an issue with our VPN client. A site visit to sn36 is planned on Tuesday to check the connection.

PG
